

Gauge	Measurement	Heater	Output	Capacity	Recorded rate
MPS System TRwS 405	Load cell	15 W	Digital	750 mm	1 min (Ma, Hauk)
Meteoservis MRW500	Load cell	350 W	Digital	900 mm	1 min (Ma, BrLa)
Sutron TPG	Load cell	200 W	Digital	914 mm	6 s (Ma)
1500 mm Geonor	Vibrating wire	60 W (Weis), N/A* (BrLa, CaCr, Ma)	Analog	1500 mm	6 s (Ma, Weis), 1 min (BrLa, CaCr)
600 mm Geonor	Vibrating wire	60 W	Analog	600 mm	6 s (Ma)
OTT Pluvio ²	Load cell	50 W	Digital	1500 mm	6 s (CARE)

* The 1500 mm Geonor gauges at Bratt's Lake, Caribou Creek, and Marshall were tested without heaters.